TITLE:

Quantitative analysis of neodymium, uranium, and palladium in nitric acid solution by reflection absorption spectrophotometry

AUTHOR(S):

Fujii, Toshiyuki; Egusa, Souichirou; Uehara, Akihiro;
Yamana, Hajimu; Morita, Yasuji

CITATION:

Fujii, Toshiyuki ...[et al]. Quantitative analysis of neodymium, uranium, and palladium in nitric acid solution by reflection absorption spectrophotometry. Journal of Radioanalytical and Nuclear Chemistry 2013, 295(3): 2059-2062

ISSUE DATE:

2013-03

URL:

<http://hdl.handle.net/2433/187114>

RIGHT:

The final publication is available at link.springer.com; この論文は出版社版ではありません。引用の際には出版社版をご確認ご利用ください。 ;
This is not the published version. Please cite only the published version.

1 Short Communication

2
3 Quantitative analysis of neodymium, uranium, and palladium in nitric acid solution by
4 reflection absorption spectrophotometry

5
6 Toshiyuki Fujii,^{1*} Souichirou Egusa,^{1,2} Akihiro Uehara,¹ Hajimu Yamana,¹ and Yasuji
7 Morita³

8
9 ¹ Division of Nuclear Engineering Science, Research Reactor Institute, Kyoto
10 University, 2-1010, Asashiro Nishi, Kumatori, Sennan, Osaka, 590-0494, Japan

11 ² Graduate School of Engineering, Kyoto University, Kyoto daigaku-Katsura,
12 Nishikyo-ku, Kyoto 615-8530, Japan

13 ³ Nuclear Science and Engineering Directorate, Japan Atomic Energy Agency,
14 Tokai-mura, Ibaraki-ken 319-1195, Japan

15
16 *Author to whom correspondence should be addressed

17 tosiyuki@rri.kyoto-u.ac.jp

18 TEL: +81-72-451-2469, FAX: +81-72-451-2634

19
20 E-mail: tosiyuki@rri.kyoto-u.ac.jp

Abstract

Quantitative analysis of Nd, U, or Pd in 3 mol dm⁻³ (M) HNO₃ was performed by reflection absorption spectrophotometry at ultraviolet-visible-near-infrared (UV/Vis/NIR) region. A sample chamber with optics for reflection measurement was designed and attached to a UV/Vis/NIR spectrophotometer by optical fibers. The reflection absorbance showed linear relations with concentrations of Nd, U, and Pd at the absorbance region less than 0.1. The quantitative analysis was found to be possible for 3 M HNO₃ solutions containing [Nd] ~ 0.2 M, [U] ~ 0.04 M, or [Pd] ~ 0.01 M.

Keywords

Absorption spectrophotometry; Reflectance; Neodymium; Uranium; Palladium; Nitric acid.

Introduction

Ultraviolet-visible-near-infrared (UV/Vis/NIR) absorption spectrophotometry is useful for quantitative analysis of complexes in solutions. For the purpose of process control in the nuclear fuel cycle, applicability of the absorption spectrophotometry is investigated [1,2]. The transmission absorption spectrophotometry with optical cells is commonly used for quantitative analysis. The method is highly precise, but basically off-line. The liquid sample is placed in an optical cell and its transmittance is measured. Flow cell is designed for online and in-line measurements of liquid stream. The liquid sample runs through an optical cell and the transmittance of the cell is measured. If the transmission system is installed to the nuclear fuel cycle, the system should equip the optical windows which contact with the radioactive solutions. Fragility of the optical windows may be problematic for avoiding a leak of the solution from the system. Without the optical windows, the transmittance of sample solution may be measured if a part of optics is directly immersed in the sample. An optical probe called "immersion probe", which consists of optical fibers and stainless steel (or hastelloy, titanium, quartz, and so on), is designed for this purpose. Direct contact of the solution with optical material is undesirable from the viewpoint of corrosion. If one considers high temperature systems like pyroprocessing or vitrification, the material damage becomes significant.

The reflection absorption spectrophotometry is known as a high-sensitive analytical method for chemical species adsorbed onto metallic surface [3]. The reflection technique is generally used in the infrared region [3]. A utilization of this technique at UV/Vis region has been reported for Langmuir monolayer [4]. A chemical species adsorbed onto the interface between dodecane and sulfuric acid has been analyzed by UV/Vis reflection absorption spectrophotometry [5]. An advantage of the

reflection technique may be that the optics can be arranged above the solution surface and do not contact with the sample solution. Optical windows are not required for the container of the sample solution. It may be utilizable for online monitoring of liquid samples treated in the nuclear fuel cycle. We report the applicability of UV/Vis reflection absorption spectrophotometry for nitric acid solutions containing Nd, U, or Pd, in which Nd and Pd are major fission product (FP) elements of rare earth elements and platinum group elements, respectively.

Experimental

All chemicals were reagent grade. $\text{Nd}(\text{NO}_3)_3 \cdot 6\text{H}_2\text{O}$ was dissolved in HNO_3 to prepare various concentrations of Nd(III) in 3 mol dm^{-3} (M) HNO_3 . Similarly, $\text{UO}_2(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$ was dissolved in 3 M HNO_3 . Pd metal was dissolved in boiling 13.3 M HNO_3 [6], and then the solution was diluted with H_2O to prepare Pd sample of 3 M HNO_3 .

The absorption spectrum was measured by using a UV/Vis/NIR spectrophotometer (JASCO, V-7200) at room temperature. A cubic sample chamber of ~200 mm on a side was designed for reflection absorption measurement, whose plane view is shown in Fig. 1 schematically. The chamber was connected with the spectrophotometer by optical fibers. An aluminum mirror was set at the focal point. The incident angle of analytical light was designed to be 5° . The deuterium and tungsten light sources were switched at 350 nm. The reflection intensity was measured in the wavelength range of 200-1000 nm at 1 nm intervals of 0.2 sec integration time each. The band widths were 1 nm (200-800 nm) and 5 nm (800-1000 nm). The detectors, secondary electron multiplier (SEM) for UV/Vis region and PbS semiconductors for NIR region, were switched at 800 nm. The obtained data was used as the blank

spectrum. Then, the aluminum mirror was replaced by the sample solution in a quartz cell, and the reflectance was measured. A metal free HNO_3 solution with the same acidity was also analyzed as reference. The transmission absorption spectrum was also measured for comparison.

Results and discussion

Neodymium possesses characteristic absorption bands in UV/Vis/NIR region [7]. The reflection intensity (R) of water and 3 M HNO_3 solution with or without Nd(III) was measured (Fig. 2a). A low signal-to-noise (S/N) ratio at around 800 nm is due to switching of detectors. The reflectance of water was measured to be $\sim 0.08\%$ relative to that of the aluminum mirror. The reflectance of the aluminum mirror is commonly certified to be $\sim 90\%$. The refractive indexes of water and quartz are 1.33 and 1.46 (Vis region, room temperature) [8]. From the Fresnel equations with the refractive indexes, for the system with 5° incident and reflection angles, 93% of analytical lights transmit the quartz cell and 0.2% of the transmitted lights should be reflected at the water surface. The R value of 0.08% means that $\sim 40\%$ of the reflected lights are successfully collected at the inlet of optical fibers directed to the detector. The absorption band at 960 nm is attributable to the vibrations of H_2O . This is assigned to the combination transition $2\nu_1 + \nu_3$, where ν_1 is the symmetric O-H stretch and ν_3 the asymmetric O-H stretch modes [9]. The absorption in UV region for 3 M HNO_3 solution is due to the electron transition of NO_3^- [10,11].

The transmission absorbance (A) is defined by the Lambert-Beer law, $A = -\log(I/I_0)$, where I_0 and I are the intensity of incident light and transmitted light, respectively. Similar to A , the reflection absorbance (A_R) is defined as,

$$A_R = -\log (R/R_0) \quad (1)$$

110

111 where R_0 is the reflection intensity measured for solvent (3 M HNO_3) in quartz cell, and
112 R is that containing solutes. The absorption spectra of Nd obtained are shown in Fig. 2b,
113 in which the transmission absorption spectrum reproduces the spectra reported for
114 nitrate systems [12-16]. The reflection absorption spectrum is similar to the
115 transmission absorption spectrum but the absorbance A_R was found to be smaller than A .
116 The A_R values in Fig 2b are shown in Fig 2c as a function of A . At the absorbance
117 region from 0 to 0.1, A_R is identical with A , while A_R is less sensitive at $A > 0.1$. This is
118 due to the detection limit of our analytical apparatus. The analytical light intensity is
119 depreciated by connecting the sample chamber with optical fibers, and under this
120 condition, the reflected light intensity of $R < 0.08\%$ is analyzed. For several absorption
121 bands positioned by arrows shown in Fig. 2b, the dependence of A_R on Nd
122 concentration was investigated. The result of A_R vs $[\text{Nd}]$ is shown in Fig. 2d. It is clear
123 that, in the region $A_R \leq 0.1$, A_R is proportional to $[\text{Nd}]$. Quantitative analysis of ~ 0.2 M
124 Nd in 3 M HNO_3 is possible by employing the calibration lines for absorption bands at
125 460 nm and 679 nm, whose coefficient of determination is better than 0.997. For
126 quantitative analysis of HNO_3 solutions with lower Nd concentration, A_R of strong
127 absorption bands like 578 nm should be useful.

128 The A and A_R values were measured for U in 3 M HNO_3 . The absorption
129 spectra obtained are shown in Fig. 3a. The transmission absorption spectrum perfectly
130 reproduced the pioneering work [17]. Uranium dissolved in HNO_3 is U(VI) and forms
131 uranyl ion, UO_2^{2+} . The electronic structure of UO_2^{2+} has been studied in detail [18,19].
132 The absorption bands possesses peaks via the stretching vibration of $\text{O}=\text{U}=\text{O}$. The

reflection absorption spectrum is similar to the transmission absorption spectrum. The A_R values are shown as a function of U concentration (Fig. 3b). Linear correlation is found in the region of $A_R < 0.1$. Quantitative analysis of ~ 0.04 M U in 3 M HNO_3 is possible by employing the calibration lines for absorption bands at 370 nm and 468 nm, whose coefficient of determination is better than 0.994. The curve of A_R for absorption bands at 415 nm can be used as an alternative of the calibration lines.

Palladium is known as one of the noble metal fission product elements in reprocessing of spent nuclear fuels. Its complexation characteristics in solvent extraction processes [20,21] are hence of interest. In our previous study, the complexation of Pd^{2+} with NO_3^- was studied in detail by transmission absorption spectrophotometry [6]. The A and A_R values of Pd in 3 M HNO_3 obtained in the present study are shown in Fig. 4a. The transmission absorption spectrum reproduced the reported spectra [6,22,23]. The reflection absorption spectrum is similar to the transmission absorption spectrum. The A_R values are shown as a function of Pd concentration (Fig. 4b). Again, the linear correlation can be obtained in the region of $A_R < 0.1$. Quantitative analysis of ~ 0.01 M Pd in 3 M HNO_3 is possible by employing the calibration line resulted from absorption intensities at the shoulder of absorption bands. In diluted [Pd] region, peak maximum can be used to prepare the calibration line.

Increasing the intensity of light source and the sensitivity of detectors improves the current system of reflection absorption spectrophotometry. Besides that, measurement of the surface of sample solution without using quartz cell should increase the reflectance light intensity. From the Fresnel equations, if we do not use the quartz cell, $\sim 2\%$ of the transmitted lights are reflected at the solution surface. This means that the reflectance light intensity increases ~ 10 times more than that with the quartz cell.

The nonflatness of liquid surface due to the surface tension and vibrations (microvibrations) under the experimental circumstance is problematic. The former shifts the focal point and the latter depresses the S/N ratio. Installing autofocus device, vibration-free system, and multiple detector may be required.

Conclusions

The UV/Vis/NIR reflection absorption technique for nitric acid solutions was developed. Quantitative analysis of Nd, U, and Pd was possible for 3 M HNO₃ solution containing ~0.2 M Nd, ~0.04 M U, or ~ 0.01 M Pd. By using A_R of stronger/weaker absorption bands, possible concentration range for quantitative analysis can be expanded. This technique may be applicable for analyzing radioactive solutions in nuclear fuel cycle.

Acknowledgment

The present study includes the result of “Development of High-level Liquid Waste Conditioning Technology for Advanced Nuclear Fuel Cycle” entrusted to “Japan Atomic Energy Agency” by the Ministry of Education, Culture, Sports, Science and Technology of Japan (MEXT).

References

- [1] Relan GR, Dubey AN, Vaidyanathan S (1995) J Radioanal Nucl Chem 204:15
- [2] Fujii T, Uda T, Fukasawa K, Uehara A, Sato N, Nagai T, Kinoshita K, Koyama T, Yamana H (2012) J Radioanal Nucl Chem DOI: 10.1007/s10967-012-2008-3
- [3] Greenler RG (1966) J Chem Phys 44:310
- [4] Okamura E, Hasegawa T, Umemura J (1995) Biophys J 69:1142

- 181 [5] Moriya Y, Hasegawa T, Hayashi K, Maruyama M, Nakata S, Ogawa N (2003)
- 182 Anal Bioanal Chem 376:374
- 183 [6] Fujii T, Egusa S, Uehara A, Kirishima A, Yamagishi I, Morita Y, Yamana H
- 184 (2011) J Radioanal Nucl Chem 290:475
- 185 [7] Yatsimirskii KB, Davidenko NK (1979) Coord Chem Rev 27:223
- 186 [8] Hynes WM (2011) Handbook of chemistry and physics, 92th edn. CRC Press,
- 187 Boca Raton
- 188 [9] Phelan MK, Barlow CH, Kelley JJ, Jinguji TM, Callis JB (1989) Anal Chem
- 189 61:1419
- 190 [10] Jones N, Thorn GD (1949) Canad J Res B 27:580
- 191 [11] Bayliss NS, Watts DW (1963) Aust J Chem 16:943
- 192 [12] Maeck WJ, Kussy ME, Rein JE (1965) Anal Chem 37:103
- 193 [13] Choppin GR, Henrie DE, Buijs K (1966) Inorg Chem 5:1743
- 194 [14] Abrahamer I, Marcus Y (1968) J Inorg Nucl Chem 30:1563
- 195 [15] Janssens-Maenhout G, Nucifora S (2007) Nucl Eng Design 237:1209
- 196 [16] Fujii T, Okude G, Uehara A, Sekimoto S, Hayashi H, Akabori M, Minato K,
- 197 Yamana H (2011) J Radioanal Nucl Chem 288:181
- 198 [17] Betts RH, Michels RK (1949) J Chem Soc S286
- 199 [18] McGlynn SP, Smith JK (1961) J Mol Phys 6:164
- 200 [19] Denning RG (1992) Struct Bonding 79:215
- 201 [20] El-Reefy SA, Daoud JA, Aly HF (1992) J Radioanal Nucl Chem 158:303
- 202 [21] Fujii T, Yamana H, Watanabe M, Moriyama H (2001) J Radioanal Nucl Chem
- 203 247:435
- 204 [22] Jørgensen CK, Parthasarathy V (1978) Acta Chem Scand A32:957
- 205 [23] Frias EC, Pitsch HK, Ly J, Poitrenaud C (1995) Talanta 42:1675
- 206

Figure captions

Fig. 1 Optical alignment of sample chamber for reflection absorption measurement.

Fig. 2 Reflection absorption spectra of Nd(III) in 3 M HNO₃. **a** Reflectance of H₂O, 3 M HNO₃, and 0.04 M Nd(III) in 3 M HNO₃. **b** Transmission absorbance A and reflection absorbance A_R for 0.04 M Nd(III) in 3 M HNO₃. Absorbances at wavelengths positioned by arrows were used in Fig. 2d. **c** A_R vs A **d** Dependence of A_R on Nd(III) concentration.

Fig. 3 Reflection absorption spectra of U(VI) in 3 M HNO₃. **a** Transmission absorbance A and reflection absorbance A_R for 0.0075 M U(VI) in 3 M HNO₃. Absorbances at wavelengths positioned by arrows were used in Fig. 3b. **b** Dependence of A_R on U(VI) concentration.

Fig. 4 Reflection absorption spectra of Pd(II) in 3 M HNO₃. **a** Transmission absorbance A and reflection absorbance A_R for 0.001 M Pd(II) in 3 M HNO₃. Absorbances at wavelengths positioned by arrows were used in Fig. 4b. **b** Dependence of A_R on Pd(II) concentration.